

# Cleaning & Sterilization Instructions



8400 Green Meadows Dr. N.  
Lewis Center, OH 43035 USA  
www.abrasive-tech.com

**EC REP** MedNet EC-REP GmbH  
Borkstrasse 10,  
48163 Muenster, Germany

<b>Manufacturer: Abrasive Technology, Inc.</b>	
<b>Devices: Two Striper® Carbopost®, Snowpost®, and Snowlight® Dental Posts</b>	
<b>WARNINGS</b>	<p><b>General</b></p> <ol style="list-style-type: none"> <li>Follow all applicable solution or equipment manufacturer's instructions.</li> <li>Use only FDA-cleared sterilizers, and sterilization accessories, per the equipment or accessory manufacturer's instruction.</li> <li><b>These devices are NOT supplied sterile.</b> Clean and sterilize prior to use, as follows.</li> </ol>
<b>Processing limitations</b>	None

<b>INSTRUCTIONS</b>	
<b>Initial treatment at the point of use</b>	Not applicable
<b>Preparation before cleaning</b>	Not applicable
<b>Cleaning: Automated</b>	Not applicable
<b>Cleaning: Manual</b>	Wipe the post using cotton soaked in ethyl or isopropyl alcohol, minimum 70%.
<b>Disinfection</b>	Not applicable
<b>Drying</b>	Dry the post with clean, pressurized air that is free of oil and water.
<b>Maintenance, Inspection and Testing</b>	Visually examine the devices for contamination. Magnification is recommended. <b>Repeat the above steps for any devices that are still contaminated.</b>
<b>Packaging</b>	Sterilization pouch
<b>Sterilization</b>	Steam sterilize, pre-vacuum, average pressure of 30psi (207 kPa), minimum of 4 minutes, 132°C, and a minimum dry time of 22 minutes.
<b>Storage</b>	Store the devices in the sterilized wrap.
<b>Additional Information</b>	None
<b>Manufacturer Contact</b>	740-548-4100

The instructions provided above have been validated by the manufacturer of the medical device as being capable of preparing a medical device for use/reuse. It remains the responsibility of the processor to ensure that the processing, as actually performed using equipment, materials and personnel in the processing facility, achieves the desired result. This requires verification and/or validation and routine monitoring of the process. **Date issued: 10/2019**